# Certificate of analysis

| | | | |
|---|---|---|---|
| **Product name:** | Testsimplets | **Formula:** | - |
| **Product code:** | 191574 | **Molecular weight** | - |
| **Lot-No:** | 39482 | **CAS No.:** | - |
| **Durability** | 02/2028 | **Storage:** | + 2 °C - 30 °C |

| Parameter | Specification | Batch values | approved |
|---|---|---|---|
| **Coating with Cresyl Violet Acetate in µg/cm²** | 1,64 – 2,44 | 1,77 – 2,41 | yes |
| **Coating with New Methylene Blue in µg/cm²** | 0,80 – 1,19 | 0,88 – 1,17 | yes |
| **Haemogram** | Clear Differentiation | As specified | yes |
| | | tested: 10.02.2026 | |

The lot values comply with the specifications.

Dr. Wolfgang Schräder
(Waldeck Production)

- This document was issued electronically and is therefore valid without signature -

Waldeck GmbH & Co. KG
Havixbecker Str. 62
48161 Münster
Amtsgericht HRA 1926

Komplementärin:
Wahlig Beteiligungs- und
Verwaltungsgesellschaft mbH
Amtsgericht Münster HRB 692

Geschäftsführer:
Wilhelm Wahlig
Ingo Gerdes

Kontakt:
Tel.: +49(0)2534/970-0
Fax: +49(0)2534/970-195
info@waldeck-ms.de
www.waldeck-ms.de