

Quality Control Certificate

Product name: Petri Dish, without vents
Item-No: 300305
Lot No: F201237W
Sterility: non sterile

Parameter	Result
Deformation by heat (8h at 80°C)	passed
Visual inspection (scratches etc.)	passed
Print on box	Complies with sample
Vents existent	no
	Inspected on 29.01.2021

Anne-Kathrin Raddatz
 (Waldeck production)

- This document was generated electronically and is valid without signature -