# Quality Control Certifcate

**Product name:** Petri Dish, without vents
**Item-No:** 300305
**Lot No:** F240337B
**Sterility**: non sterile

| Parameter | Result |
|---|---|
| **Deformation by heat (8h at 80°C)** | passed |
| **Visual inspection (scratches etc.)** | passed |
| **Print on box** | Complies with sample |
| **Vents existent** | no |
| | |
| | Inspected on 19.04.2024 |

Anne-Kathrin Raddatz
(Waldeck production)

- This document was generated electronically and is valid without signature -

Waldeck GmbH & Co. KG
Havixbecker Str. 62
48161 Münster/Germany
County Court Münster
HRA 1926

Unlimited Partner:
Wahlig Beteiligungs- und
Verwaltungsgesellschaft mbH
County Court Münster
HRB 692

Managing Directors:
Wilhelm Wahlig
Ingo Gerdes

Contact:
Phone: +49(0)180/2 24 76 62
Fax: +49(0)180/1 24 76 62
info@waldeck-ms.de
www.waldeck-ms.de