

## Quality Control Certificate

**Product name:** Petri Dish, with vents  
**Item-No:** 300306  
**Lot No:** F140615D  
**Sterility:** non sterile

Parameter	Result
<b>Deformation by heat (8h at 80°C)</b>	passed
<b>Visual inspection (scratches etc.)</b>	passed
<b>Print on box</b>	Complies with sample
<b>Vents existent</b>	yes
	Inspected on 08.07.2014

Dr. Wolfgang Schröder  
 (production manager)

- This document was generated electronically and is valid without signature -