

**Quality Control Certificate**

**Product name:** Petri Dish, with vents  
**Item-No:** 300306  
**Lot No:** F180133A  
**Sterility:** non sterile

<b>Parameter</b>	<b>Result</b>
<b>Deformation by heat (8h at 80°C)</b>	passed
<b>Visual inspection (scratches etc.)</b>	passed
<b>Print on box</b>	Complies with sample
<b>Vents existent</b>	yes
	Inspected on 24.01.2018

Anne-Kathrin Raddatz  
(Waldeck production)

- This document was generated electronically and is valid without signature -