Quality Control Certifcate

Product name: Petri Dish, with vents
Item-No: 300306
Lot No: F220934F
Sterility: non sterile

Parameter	Result
Deformation by heat (8h at 80°C)	passed
Visual inspection (scratches etc.)	passed
Print on box	Complies with sample
Vents existent	yes
	Inspected on 10.10.2022

Anne-Kathrin Raddatz
(Waldeck Production)

- This document was generated electronically and is valid without signature -

Waldeck GmbH & Co. KG
Havixbecker Str. 62
48161 Münster/Germany
County Court Münster
HRA 1926

Unlimited Partner:
Wahlig Beteiligungs- und
Verwaltungsgesellschaft mbH
County Court Münster
HRB 692

Managing Directors:
Wilhelm Wahlig
Ingo Gerdes

Contact:
Phone:+49(0) 2534 / 970-0
Fax: +49(0) 2534 / 970-195
info@waldeck-ms.de
www.waldeck-ms.de