Certificate of analysis

Product name:	Testsimplets	**Formula:**	-
Product code:	191574	**Molecular weight**	-
Lot-No:	33972	**CAS No.:**	-
Durability	11/2026	**Storage:**	+ 2 °C - 30 °C

Parameter	Specification	Batch values	approved
Coating with Cresyl Violet Acetate in µg/cm²	1,64 – 2,56	1,85 – 2,41	yes
Coating with New Methylene Blue in µg/cm²	0,80 – 1,19	0,80 – 1,03	yes
Haemogram	Clear Differentiation	As specified	yes
		tested: 19.11.2024	

The CoA was corrected on 10-03-2025 by W. Schräder: wrong Durability

The lot values comply with the specifications.

Dr. Wolfgang Schräder
(Waldeck Production)

- This document was issued electronically and is therefore valid without signature -

Waldeck GmbH & Co. KG
Havixbecker Str. 62
48161 Münster
Amtsgericht HRA 1926

Komplementärin:
Wahlig Beteiligungs- und
Verwaltungsgesellschaft mbH
Amtsgericht Münster HRB 692

Geschäftsführer:
Wilhelm Wahlig
Ingo Gerdes

Kontakt:
Tel.: +49(0)2534/970-0
Fax: +49(0)2534/97 01 95
info@waldeck-ms.de
www.waldeck-ms.de