Certificate of analysis

Product name:	Testsimplets	**Formula:**	-
Product code:	191574	**Molecular weight**	-
Lot-No:	20546	**CAS No.:**	-
Durability	01/2024	**Storage:**	+ 2 °C - 30 °C

Parameter	Specification	Batch values	approved
Coating with Crystal Violet Acetate in µg/cm²	1,64 – 2,56	1,86 – 2,27	yes
Coating with New Methylene Blue in µg/cm²	0,80 – 1,19	0,81 – 0,95	yes
Haemogram	Clear Differentiation	As specified	yes
		tested: 25.01.2022	

Correction on 30.09.2022: storage temperature was corrected to "+2°C to 30°C"

The lot values comply with the specifications.

Dr. Wolfgang Schräder
(Waldeck Production)

- This document was issued electronically and is therefore valid without signature -

Waldeck GmbH & Co. KG
Havixbecker Str. 62
48161 Münster
Amtsgericht HRA 1926

Komplementärin:
Wahlig Beteiligungs- und
Verwaltungsgesellschaft mbH
Amtsgericht Münster HRB 692

Geschäftsführer:
Wilhelm Wahlig

Kontakt:
Tel.: +49(0)2534/970-0
Fax: +49(0)2534/97 01 95
info@waldeck-ms.de
www.waldeck-ms.de